


Instruction for use rigid endoscopes

Scope of validity:

520170FX, 520190FX, 520970FX, 521090FX, 521170FX, 521190FX, 521670FX, 521671FX, 521672FX, 522090FX, 700900FX, 700912FX, 700930FX, 700970FX, 701300FX, 701330FX, 716000FX, 716030FX, 716100FX, 716130FX, 716200FX, 716230FX, 717100FX, 717130FX, 717170FX, 718100FX, 718130FX, 718200FX, 718230FX, 718300FX, 718330FX, 718400FX, 718430FX, 722100FX, 722130FX, 722170FX, 728100FX, 728130FX, 728131FX, 728170FX, 728200FX, 728230FX, 728270FX, 728500FX, 728530FX, 728570FX, 728600FX, 728630FX, 728645FX, 728670FX, 728830FX, 730100FX, 730130FX, 730145FX, 730170FX, 740210FX, 750100FX, 750110FX, 750130FX, 750131FX, 750145FX, 750146FX, 750170FX, 750171FX, 750190FX, 750270FX, 750300FX, 750330FX, 750345FX, 750370FX, 750400FX, 750430FX, 750445FX, 750470FX, 750600FX, 806270FX.

1. USED SYMBOLS			
	Manufacturer		Consult Instructions
	Caution		Article number
	Production batch, batch		Non steril
	CE- labeling		
2. PURPOSE OF THIS INSTRUCTION			
<p>This document describes the correct handling and function of the rigid endoscope, as well as recommended processing methods. This document may not be used to carry out endoscopic examinations or surgeries, nor may it be used for training purposes.</p> <p>The respective current version of this document can be requested from FENTEX medical GmbH. If you as the user of this endoscope believe that you require more detailed information regarding the product's use and maintenance, please contact your representative.</p>			
3. INTENDED USE			
<p>Rigid medical endoscopes are used to visualize body cavities. Each endoscope was developed for diagnostic and surgical procedures in one of the following fields of application:</p> <ul style="list-style-type: none"> • Sinuscope: sinusopic procedures • Otoscope: otoscopic procedures • Laryngopharyngoscope: Laryngopharyngoscopic procedures <p>For the benefit and safety of patients, physicians must select a method which they consider suitable based on their experience. If you as the user of this endoscope believe that you require more detailed information regarding the product's use and maintenance, please contact your representative.</p>			
4. CONTRAINDICATIONS			
<p>BSE - Bovine spongiform encephalopathy, so-called "mad cow disease" CJK - Creutzfeldt-Jakob disease TSE - Transmissible spongiform encephalopathy VCJD - Variant Creutzfeldt-Jakob disease</p> <p>Currently, no contraindications are known that are directly related to the endoscopes.</p>			
5. TESTING, HANDLING AND MAINTENANCE			
5.1 Construction			
		<p>1 Distal end 2 Sheath 3 Main part 4 Eyepiece 5 Proximal end 6 Irradiation surface of the light cable 7 Connection for light cable type ACMI 8 Adapter for light cable type Wolf 9 Adapter for light cable type Storz /Olympus</p>	
5.2 Markings			
<ul style="list-style-type: none"> • Article number • Serial number • CE mark • For autoclavable endoscopes: Marking "autoclavable" 			



- Specification of the direction of view
- Marking GERMANY
- Marking FENTEX MEDICAL
- For high resolution optics: Marking HD

5.3 Available design and sizes

The endoscopes are available in the following designs and sizes:

- Straight endoscopes
- Sheath diameter 1.9 - 10 mm

5.4 Combinable products

You can combine the endoscopes with existing camera systems and with light cables and instruments from FENTEX medical.

6. SAFETY INSTRUCTION

The endoscope may only be used by trained medical professionals in medical facilities.

- After delivery, inspect the endoscope for completeness and damage.
- Read, observe and keep the instructions for use.
- Use the endoscope only as intended, see chapter 3 "Intended use"

For storage, transport and processing, ensure that the endoscope is not subjected to mechanical strain, particularly to prevent damage to the sensitive lens system.

WARNING

Risk of infection to the patient or medical professionals!

The endoscopes are delivered non-sterile as reusable products.

The state of the art and national laws require the observance of validated processes.

In general, users are responsible for the validation of their processes.

- **Ensure that the processing, material and personnel are suitable for achieving the results necessary.**
- **Observe any valid local operator regulations for all manual cleaning and drying processes.**
- **Clean / disinfect and sterilize the endoscope prior to initial use as well as each subsequent use of the endoscope.**
- **Bring the endoscope to the decontamination area after use. Observe valid protective measures to prevent contaminating the environment.**

WARNING

Risk of burns!

The optical fibres emit high-energy light at the distal end of the endoscope. This can cause the temperature of the body tissue to rise to 41 °C.

- **Avoid direct contact of the distal end with body tissue or flammable materials as it can cause burns.**
- **Reduce the light intensity of the cold light source when working near body tissue or flammable materials.**

WARNING

Risk of injury due to faulty endoscopes!

- **Carry out visual inspection and function check prior to each use.**
- **Only use endoscopes which are in perfect condition.**

Endoscopes from FENTEX medical are precision medical instruments, and handling them requires great care.

- Inspect the endoscope for damage prior to and after use.
- If the endoscope is damaged, discontinue use and contact the manufacturer.
- Do not subject the endoscope to impact.
- Put the endoscope down carefully.
- Hold endoscope only by the eyepiece / main part and not by the sheath.
- Do not bend the sheath.
- Do not bend the body after inserting the endoscope into the body. A piece broken off the endoscope can become lodged in the soft tissue or no longer appear in the endoscope's field of vision and thus remain in the body.
- Transport endoscopes individually and store them safely by using a wire basket or container.

7. PREPARATION FOR USE

7.1 Visual inspection and function check

WARNING

Risk of injury due to faulty endoscopes!

Carry out a visual inspection and function check prior to initial use as well as each additional use.

- **Only use endoscopes which are in perfect condition.**



- Clean / disinfect and sterilize the endoscope prior to initial use as well as each additional use of the endoscope. Contaminants on the irradiation surface of the illumination fibres can burn in during use, which impacts image quality.
- Ensure that the proximal end of the endoscope is dry to prevent the endoscope from fogging up during the examination / procedure.
- Ensure that no parts are missing or loose.
- Ensure that there are no residual cleaning agents or disinfectants on the endoscope.
- Inspect the entire endoscope, particularly the sheath, for contaminants and damage of any type, such as dents, scratches, cracks, bending and sharp edges.
- Inspect distal end, proximal end and irradiation surface of the illumination fibres for contamination and scratches. Make contaminants and scratches visible using light reflexes. Hold the connection of the optical fibres against the light and inspect whether the optical fibres illuminate evenly at the distal end.
- Check image quality: The image may not be blurry, clouded or dark. If necessary, remove deposits on the optical end surface using polishing paste provided, see "Removing deposits from optical end surfaces"
- For endoscopes with locking device: Inspect between the sheath and the main part for contaminants and damage to ensure a fixed and secure connection.

7.2 Provisioning

- Clean / disinfect and sterilize the endoscope prior to initial use as well as each additional use of the endoscope, see "Processing"
- Ensure that the proximal end of the endoscope is dry to prevent the endoscope from fogging up during the examination/procedure.
- If necessary mount adapter for light cable, see "Assembly"
- Mount light cable (see manufacturer's specifications).
- If required, adapt the camera (see manufacturer's specifications).



8. INSTRUCTIONS FOR REPROCESSING ACCORDING TO DIN 17664

8.1 Safe storage and transport

If possible, reprocessing endoscopes immediately after use is recommended. Endoscope containers and trays are reusable. Trays must be inspected for visible contamination and cleaned prior to use. They can be cleaned manually or in an automatic cleaning unit using a cleaning agent.

- Always store endoscope securely and transport it to processing in a closed container to prevent damage to the endoscope and contamination of the environment.

8.2 Cleaning and disinfect

Manual cleaning / pre-cleaning and chemical disinfection

Do not use fixating cleaning agents or hot water (>40 °C) as it can cause fixation of the contaminants and jeopardise successful cleaning. Do not scratch contaminants off with hard objects as this can cause damage to the optical end surfaces. Do not clean endoscope in an ultrasonic bath.

- Existing adapters are dismantled.
- For laryngoscopes: Handle is dismantled.
- Remove coarse contamination from the endoscope. Clean the endoscope with a soft brush under cold tap water until all visible contaminants have been removed.

The endoscopes have material compatibility with the Steris™ System 1 process.

- Disinfect endoscope. In doing so, observe the specifications of the disinfectant solution manufacturer regarding temperature, concentration and application time.

Non-compliance with the manufacturer's specifications can result in damage to the endoscope.

- Rinse endoscope with running water.
- Dry endoscope with a soft cloth.
- Carry out visual inspection, function check and servicing, see "Testing, handling and maintenance"

Machine cleaning and thermal disinfection

The rigid endoscopes from FENTEX medical are suitable for prevalent machine methods of cleaning and thermal disinfection. In doing so, use gentle cycles for rigid endoscopes and suitable cleaning agents and disinfectants. The instructions of the machine, cleaning agent and disinfectant manufacturers must be observed. The cleaning and disinfectant result must be confirmed by the machine, cleaning agent and disinfectant manufacturers in cooperation with the user.

The following methods have been validated by FENTEX medical for the following rigid endoscopes:

- Fix the endoscope to the loading rack in such a way that damage is prevented during cleaning.

The following materials and machines were used for the validation:

Cleaning agents	-Alkaline: Neodisher FA; Dr. Weigert; Hamburg -Enzymatic: Endozime, Ruhof;
Neutraliser	- Neodisher Z; Dr. Weigert, Hamburg
Cleaning and disinfecting unit	- Miele G 7736 CD
Loading racks	- Loading rack E 327-06 - MIC rack E 450


- Start cleaning process:
 - Pre-rinse with cold water for 1 minute, drain
 - Pre-rinse with cold water for 3 minutes, drain
 - Clean with 0.5% alkaline cleaning agent for 5 minutes at 55 °C or with 0.5% enzymatic cleaning agent at 45 °C, drain
 - Neutralise for 3 minutes with warm tap water (<40 °C) and neutralizer, drain
 - Intermediate rinse for 2 minutes with warm tap water (<40 °C), drain
- Carry out machine thermal disinfection considering the national requirements regarding the Ao value (see DIN EN ISO 15883)
- Ensure that the exteriors of the endoscope are dry. If necessary, dry with a soft cloth.
- Carry out visual inspection, function check and servicing, see "Testing, handling and maintenance"

Removing deposits from optical end surfaces

If deposits are found when checking the image quality, they can be removed with the provided polishing paste as follows:

Only clean with polishing paste if the image which you see through the endoscope is cloudy and blurry

- Apply polishing paste to a clean cotton swab.
- For large end surfaces: press cotton swab lightly on the end surface to be cleaned and rub it over the glass.
- For small end surfaces: place cotton swab lightly on the end surface to be cleaned and turn it.



- Clean all optical end surfaces with warm water and mild detergent to remove all polishing paste residue.
- Rinse optical end surfaces under running water.
- Dry optical end surfaces with a soft cloth.
- Clean / disinfect and sterilize the endoscope.
- Carry out visual inspection. If the deposits were not removed: send in endoscope for repair.

8.3 Sterilization

Prior to each sterilization, rigid endoscopes must be cleaned and disinfected according to the methods in these instructions for use.

- Sterilize endoscopes in suitable packaging to prevent subsequent contamination.

a) Steam sterilization (autoclaving)

In general, users are responsible for the validation of their processes.

Only endoscopes which are marked with the writing "autoclavable" are intended for autoclaving. The permissible processing methods are explained in the instructions at hand.

- When selecting the processing method, observe the valid national hygienic regulations and local provisions for hospital hygiene.

Comply with specified process parameters. The specified parameters have been validated to ensure the sterility of the endoscopes. Deviating process parameters could damage the endoscope. In this case, the guarantee and warranty shall become void.

Autoclavable endoscopes can be sterilized with the French cycle (134 C, 18 minutes, 3.1 bar (absolute) without restrictions regarding material compatibility.

- Existing adapters are dismantled.
- Sterilize endoscopes.
- When the sterilization process has ended, allow the endoscopes to cool gradually to room temperature.

b) Fractionated pre-vacuum method

The following process has been validated:

Temperature	132 – 137°C (270-278°F)
-------------	-------------------------



Time	At least 3 minutes (5 minutes are recommended)
Configuration	Double packed in sterilization bags
Drying	At least 10 minutes

c) Gravitation method

The endoscopes have material compatibility with the gravitation method for a hold time of 15 minutes.

d) Hydrogen peroxide sterilization (STERRAD® method)

The endoscopes can be sterilized with the following STERRAD systems:

- STERRAD 100S
- STERRAD NX
- STERRAD 100NX
 - Observe specifications of the manufacturer (ASP) regarding the corresponding method.

e) Ethylene oxide sterilization

The endoscopes are material compatible with Ethylene oxide sterilization

8.4 Special precautions: Pathogens of Transmissible Spongiform Encephalopathy

A comprehensive explanation of the necessary preventative measures with regard to agents of Transmissible Spongiform Encephalopathy (TSE) would go beyond the scope of this document.

It is assumed that pathogens of the Creutzfeldt Jakob Disease cannot be killed using normal disinfection and sterilization processes. Therefore, the standard methods for decontamination and sterilization are not sufficient if there is a risk of transferring Creutzfeldt Jakob Disease.


In general, only tissue with a low potential of TSE infection comes into contact with surgical instruments. In spite of this, special preventative measures must be taken for instruments which are used to treat patients with a known or suspected infection of TSE, as well as for patients at risk.

8.5 Processing restrictions

Repeated processing has only minimal effect on the endoscopes. The service life of the units is usually determined by wear and damage. The endoscope can be damaged if the manufacturer's specifications are not observed.

Do not clean endoscope in an ultrasonic bath.

9. ASSEMBLY



- 1 = Connection for light cable type ACMI
- 2 = Adapter type Wolf
- 3 = Adapter type Storz/ Olympus

- If necessary mount adapter for light cable
 - Ensure that the irradiation surface of the light cable is clean.
 - Mount light cable (see manufacturer's specifications).
- If required, adapt the camera (see manufacturer's specifications).

10. DISASSEMBLY



WARNING!

Risk of burns!

Allow the light cable to cool sufficiently before removing it. The ends can become very hot and cause serious burns.

- Remove light cable.
- Do not remove the eyepiece because otherwise the endoscope will be damaged.

11. STORAGE

Unsterile metal units must be stored in a clean, dry environment. The storage time of unsterile units is not limited; the units are made of a non-degradable material which maintains its stability when stored under the recommended conditions.

As long as endoscopes are stored unsterile in the original packaging, the following storage conditions apply:

- Temperature: -10 °C to +40 °C
- Humidity: 10% to 90%

- Avoid direct sunlight.
- Store endoscope either in the original packaging or in a wire basket/container.
- Ensure that the endoscope is stored securely.

Observe the respective valid national provisions when storing in a sterile condition.



12. SERVICE AND MAINTENANCE

FENTEX medical does not supply original parts to independent workshops or other endoscope manufacturers. Thus only FENTEX medical is in a position to carry out repairs using original parts. The original technical specifications and the operational safety of the endoscope can only be guaranteed by using original parts. The warranty for FENTEX medical products shall become void if repairs are carried out by a workshop not authorized by FENTEX medical. In this case FENTEX medical is also no longer responsible for the technical specifications or safety of the product.

- Have the endoscopes repaired by FENTEX medical only. For service, send the defective endoscope to the address of the sales partner.
- Clean, disinfect and sterilize the endoscope thoroughly prior to returning it for repair.
- Ideally, send in the endoscope in its original packaging. If this is not possible, package the endoscope to secure it for transport.

FENTEX medical is not liable for damage resulting from improper shipping.

13. ACCESSOIRES / SPARE PARTS

Adapter:

711173FX Adapter K. Storz for Standard Scopes

711177FX Adapter R.Wolf for Standard Scopes

14. DISPOSAL

- Observe country-specific regulations and laws for the disposal of medical products.
- Alternative: Return the product to the manufacturer (pls. see chapter 12 Service and Maintenance)